

QuickVue In-Line® Strep A Quiz

Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the QuickVue In-Line Strep A test procedure. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current package insert instructions and/or procedure card included in the kit.

1. How often does the manufacturer recommend that external controls be run on the kit?

- a) Once every 20 tests
- b) Once per each new shipment of kits
- c) Every first sample tested each day
- d) Once per each untrained operator, once for each new shipment of kits-provided that each different lot in a shipment is tested

2. At what temperature do the kits need to be stored?

- a) 2-8°C (refrigerated)
- b) 15-30°C (room temp.)
- c) 2-30°C (refrigerated or room temp.)
- d) -20°C (frozen)

3. For proper test performance, which swab should be used when collecting a throat swab specimen?

- a) The swabs provided in the kit. They are specifically engineered with an enlarged swab tip that fits snugly into the testing device well.
- b) Any type of throat swabs
- c) Any Dacron swabs
- d) Cotton tipped swabs

4. What is the read time for this kit?

- a) 5 minutes
- b) 6 minutes
- c) 10 minutes
- d) 20 minutes

5. Can the test be read after the read time?

- a) No
- b) Yes, up to 10 minutes after the read time
- c) Yes, up to 5 minutes after the read time
- d) Yes, any time after the read time

6. How long after breaking the Extraction Reagent Delivery Device is it still good?

- a) 8 hours
- b) 48 hours
- c) Immediately after crushing
- d) 24 hours

7. How long should you wait before adjusting the swab to allow for flow if liquid has not moved across the result window?

- a) Don't need to wait
- b) 5 minutes
- c) 1 minute
- d) 6 minutes

8. How much reagent should be added to the sample well after inserting the swab?

- a) All of the reagent
- b) 8 drops
- c) Fill to the top rim of the well
- d) 10 drops

9. What procedure should be followed for Proficiency Testing?

- a) The same procedure as for a patient sample
- b) There is no procedure to follow
- c) Proficiency samples do not need to be tested
- d) There is a separate procedure to follow as stated in the Package Insert.

10. If culture is desired after throat swab is collected, what is the process to follow?

- a) Lightly streak the swab on a 5% sheep blood agar plate before using swab in the QuickVue In-Line test or obtain dual throat swab specimens
- b) There is no process to follow
- c) Culture is not necessary
- d) Lightly streak the swab on a 5% sheep blood agar plate after using swab in the QuickVue In-Line Strep A test

QuickVue In-Line® Strep A Quiz

Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the QuickVue In-Line Strep A test procedure. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current package insert instructions and/or procedure card included in the kit.

1. How often does the manufacturer recommend that external controls be run on the kit?

- a) Once every 20 tests
- b) Once per each new shipment of kits
- c) Every first sample tested each day
- d) Once per each untrained operator, once for each new shipment of kits-provided that each different lot in a shipment is tested

2. At what temperature do the kits need to be stored?

- a) 2-8°C (refrigerated)
- b) 15-30°C (room temp.)
- c) 2-30°C (refrigerated or room temp.)
- d) -20°C (frozen)

3. For proper test performance, which swab should be used when collecting a throat swab specimen?

- a) The swabs provided in the kit. They are specifically engineered with an enlarged swab tip that fits snugly into the testing device well.
- b) Any type of throat swabs
- c) Any Dacron swabs
- d) Cotton tipped swabs

4. What is the read time for this kit?

- a) 5 minutes
- b) 6 minutes
- c) 10 minutes
- d) 20 minutes

5. Can the test be read after the read time?

- a) No
- b) Yes, up to 10 minutes after the read time
- c) Yes, up to 5 minutes after the read time
- d) Yes, any time after the read time

6. How long after breaking the Extraction Reagent Delivery Device is it still good?

- a) 8 hours
- b) 48 hours
- c) Immediately after crushing
- d) 24 hours

7. How long should you wait before adjusting the swab to allow for flow if liquid has not moved across the result window?

- a) Don't need to wait
- b) 5 minutes
- c) 1 minute
- d) 6 minutes

8. How much reagent should be added to the sample well after inserting the swab?

- a) All of the reagent
- b) 8 drops
- c) Fill to the top rim of the well
- d) 10 drops

9. What procedure should be followed for Proficiency Testing?

- a) The same procedure as for a patient sample
- b) There is no procedure to follow
- c) Proficiency samples do not need to be tested
- d) There is a separate procedure to follow as stated in the Package Insert.

10. If culture is desired after throat swab is collected, what is the process to follow?

- a) Lightly streak the swab on a 5% sheep blood agar plate before using swab in the QuickVue In-Line test or obtain dual throat swab specimens
- b) There is no process to follow
- c) Culture is not necessary
- d) Lightly streak the swab on a 5% sheep blood agar plate after using swab in the QuickVue In-Line Strep A test