

MICROVUE™

Bone Health

Technical Data Sheet

MicroVueOsteocalcin

Sold in the U.S. for **Research Use Only**. Not for use in diagnostic procedures.

MicroVue Osteocalcin provides a quantitative measure of intact osteocalcin (OC) in serum as a biochemical indicator of bone turnover.

Format

- ELISA
- 96-well microplate with reagents sufficient to test 40 samples in duplicate.
- Sample type: serum.
- Controls: High, Low.

Species Reactivity

- Rabbit, Pig, Sheep, Goat, Cow, Horse, Cynomolgous Macaque, Human.
- Human cell cultures.
- No recognition of Rat, Mouse, Dog.

Specimen

- Serum, processed and frozen at $\leq -20^{\circ}\text{C}$ within 4 hours of collection.
- Do not use collection tubes containing EDTA.

Assay Steps

- Add 25 μL Standards, Controls, and Samples.
- Add 125 μL anti-Osteocalcin antibody.
- Incubate 2 hours \pm 10 minutes at room temperature.
- Wash 3 times with 1X Wash Buffer.
- Add 150 μL reconstituted Enzyme Conjugate.
- Incubate 60 \pm 5 minutes at room temperature.
- Wash 3 times with 1X Wash Buffer.
- Add 150 μL Working Substrate Solution.
- Incubate 35-40 minutes at room temperature.
- Add 50 μL Stop Solution.
- Read OD at 405 nm.
- Data reduction: 4 parameter.

Performance of Test

- Method:** Competitive
- Analyte:** Intact osteocalcin
- Specimen:** Serum
- Specimen Volume:** 25 μL
- Limit of Detection:** 0.45 ng/ μL
- Lower Limit of Quantification:** 2 ng/ μL
- Higher Limit of Quantification:** 32 ng/ μL
- Precision CVs within-run:** 4.8-10.0%
- Precision CVs between-run:** 4.8-9.8%
- Incubation Time:** 3.6 Hours
- Cross Reactivity:** Intact OC 100%
Reduced OC N.D.
N-, C- peptides N.D.

TO ORDER: Call Quidel Customer service at 1.800.524.6318 (USA only), 1.408.616.4301 or fax 1.408.616.4310.

MicroVue Osteocalcin Cat. #8002

Also Available

Osteocalcin Controls Cat. #4819
(set of 4)



RESEARCH TO RAPIDS™

Worldwide Headquarters
10165 McKellar Court
San Diego, CA 92121 USA
www.quidel.com

For more information on our products and other Quidel locations, visit our website.
800.524.6318 • 408.616.4301 • 408.616.4311 (fax)

0025BH1001D-6 (01/10)