

Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the QuickVue hCG Urine Test procedure. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current Package Insert instructions and/or Procedure Card included in the kit.

---

**1. How often does the manufacturer recommend that external controls be run on the kit?**

- |                                       |  |
|---------------------------------------|--|
| a) Once every 20 tests                | d) Once for each new lot or shipment of test materials, and as otherwise required by your laboratory's standard quality control procedures |
| b) Once per each new shipment of kits |  |
| c) Every first sample tested each day |  |

---

**2. At what temperature do the kits need to be stored?**

- |                                    |   |
|------------------------------------|---|
| a) 2°C to 8°C (refrigerated)       | c) 2°C to 30°C (refrigerated or room temperature) |
| b) 15°C to 30°C (room temperature) | d) -20°C (frozen)                                 |

---

**3. Which external quality controls are recommended to be used with this kit?**

- |  |  |
|--|--|
| a) Known patient sample  | c) Any manufacturer's external quality controls      |
| b) External quality controls are not recommended to be performed | d) hCG Control Set (Cat. #00272) – (sold separately) |

---

**4. What is the read time for this kit?**

- |              |               |
|--------------|---------------|
| a) 3 minutes | c) 10 minutes |
| b) 5 minutes | d) 20 minutes |

---

**5. Can the test be read after the read time?**

- |  |   |
|--|---|
| a) No  | c) Yes, up to 5 minutes after the read time |
| b) Yes, up to 10 minutes after the read time | d) Yes, any time after the read time        |

---

**6. How long can the urine specimen be stored at room temperature before being used to test?**

- |  |             |
|--|-------------|
| a) The specimen must be used immediately | c) 8 hours  |
| b) 48 hours                              | d) 24 hours |

---

**7. What is the detection level of this test?**

- |              |              |
|--------------|--------------|
| a) 50 mIU/mL | c) 25 mIU/mL |
| b) 20 mIU/mL | d) 10 mIU/mL |

---

**8. How many drops of urine are needed to perform the test?**

- |                               |             |
|-------------------------------|-------------|
| a) 5 drops                    | c) 3 drops  |
| b) All of the urine is needed | d) 10 drops |

---

**9. When should the Test Cassette be removed from the foil pouch?**

- |                          |                       |
|--------------------------|-----------------------|
| a) 30 minutes before use | c) 2 hours before use |
| b) 1 hour before use     | d) Just before use    |

---

**10. What is the "gold standard" test in detecting pregnancy?**

- |                                     |                       |
|-------------------------------------|-----------------------|
| a) Beta hCG quantitative test       | c) Any rapid hCG test |
| b) There is no "gold standard" test | d) Culture            |

Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the QuickVue hCG Urine Test procedure. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current Package Insert instructions and/or Procedure Card included in the kit.

**1. How often does the manufacturer recommend that external controls be run on the kit?**

- a) Once every 20 tests
- b) Once per each new shipment of kits
- c) Every first sample tested each day
- d)  Once for each new lot or shipment of test materials, and as otherwise required by your laboratory's standard quality control procedures

**2. At what temperature do the kits need to be stored?**

- a) 2°C to 8°C (refrigerated)
- b)  15°C to 30°C (room temperature)
- c) 2°C to 30°C (refrigerated or room temperature)
- d) -20°C (frozen)

**3. Which external quality controls are recommended to be used with this kit?**

- a) Known patient sample
- b) External quality controls are not recommended to be performed
- c) Any manufacturer's external quality controls
- d)  hCG Control Set (Cat. #00272) – (sold separately)

**4. What is the read time for this kit?**

- a)  3 minutes
- b) 5 minutes
- c) 10 minutes
- d) 20 minutes

**5. Can the test be read after the read time?**

- a)  No
- b) Yes, up to 10 minutes after the read time
- c) Yes, up to 5 minutes after the read time
- d) Yes, any time after the read time

**6. How long can the urine specimen be stored at room temperature before being used to test?**

- a) The specimen must be used immediately
- b) 48 hours
- c)  8 hours
- d) 24 hours

**7. What is the detection level of this test?**

- a) 50 mIU/mL
- b) 20 mIU/mL
- c)  25 mIU/mL
- d) 10 mIU/mL

**8. How many drops of urine are needed to perform the test?**

- a) 5 drops
- b) All of the urine is needed
- c)  3 drops
- d) 10 drops

**9. When should the Test Cassette be removed from the foil pouch?**

- a) 30 minutes before use
- b) 1 hour before use
- c) 2 hours before use
- d)  Just before use

**10. What is the "gold standard" test in detecting pregnancy?**

- a)  Beta hCG quantitative test
- b) There is no "gold standard" test
- c) Any rapid hCG test
- d) Culture