



### Supplemental Information and Frequently Asked Questions

#### ***Supplemental Information***

**Do you have a proficiency recommendation to validate the MicroVue DPD EIA?**

At this time, the DPD assay is not covered by any of the commercial proficiency programs. Quidel recommends the laboratory use the accepted alternative method of split testing every 6 months with a partner laboratory also testing for DPD. Please check back with us for any future developments.

**What is the CMS suggested CPT code and National Limit amount for the MicroVue DPD EIA?**

The suggested\* CPT code is 82523 (Collagen Cross-links). The Medicare National Limit amount\*\* for Urine is \$25.43. For reimbursement information and support, please visit our website at [quidel.com](http://quidel.com) or contact Quidel directly at **800.874.1517 Option 2**, or via e-mail at [technicalsupport@quidel.com](mailto:technicalsupport@quidel.com).

**How many samples can be tested with each kit?**

Each kit contains enough reagents for testing approximately 40 specimens plus MicroVue DPD EIA Standards and Controls in duplicate.

**Can I purchase reagents individually?**

The DPD assay is intended to be sold as a complete kit. Please contact Quidel Technical Support at **800.874.1517 Option 2**, or via e-mail at [technicalsupport@quidel.com](mailto:technicalsupport@quidel.com) for availability of individual reagents.

**What is the proper wash technique for this assay?**

Quidel strongly recommends use of an automated plate washer or a wash bottle apparatus. Plates should be washed with a method validated with the kit. Please visit Quidel's website for more information, including a Technical Bulletin with additional information regarding proper wash technique for Bone Health assays: [http://www.quidel.com/sites/quidel.com/files/product/documents/microvue\\_bh\\_wash\\_technique\\_tb\\_2.pdf](http://www.quidel.com/sites/quidel.com/files/product/documents/microvue_bh_wash_technique_tb_2.pdf)

**When interpreting results, do I need to account for the 1:10 sample dilution factor?**

Since all Standards, Controls, and samples are diluted 1:10 with the Assay Buffer Solution included in the kit, the 10-fold sample dilution does not need to be taken into account for results interpretation. However, if a sample is diluted in any ratio other than 1:10, then the dilution will need to be correctly accounted for. For example, if a sample is diluted 1:20 instead of 1:10, a dilution factor of 2 would need to be accounted for in this particular sample.

## ***Frequently Asked Questions***

### **What is the clinical significance of DPD measurement?**

DPD (deoxypyridinoline) crosslinks can be measured as an indicator of bone resorption. This measurement can be used in monitoring bone resorption changes in postmenopausal women or in individuals with osteoporosis. For an expanded description of this significance, please refer to page 1 of the MicroVue DPD EIA Package Insert:  
[http://www.quidel.com/sites/quidel.com/files/product/documents/8007\\_microvue\\_dpd\\_english\\_0.pdf](http://www.quidel.com/sites/quidel.com/files/product/documents/8007_microvue_dpd_english_0.pdf)

### **What is Quidel's quality control recommendation for this assay?**

Quidel recommends that positive and negative controls be included in each assay. More information regarding Quality Control recommendations can be found on page 4 of the MicroVue DPD EIA Package Insert:  
[http://www.quidel.com/sites/quidel.com/files/product/documents/8007\\_microvue\\_dpd\\_english\\_0.pdf](http://www.quidel.com/sites/quidel.com/files/product/documents/8007_microvue_dpd_english_0.pdf)

### **Can I use incubation times and/or temperatures that are different than what are listed in the protocol?**

Using incubation times and/or temperatures that differ from what is listed in the Package Insert may give erroneous results, and is therefore discouraged. Please refer to the ASSAY PROCEDURE section on pages 3-4 of the MicroVue DPD EIA Package Insert:  
[http://www.quidel.com/sites/quidel.com/files/product/documents/8007\\_microvue\\_dpd\\_english\\_0.pdf](http://www.quidel.com/sites/quidel.com/files/product/documents/8007_microvue_dpd_english_0.pdf)

### **What is the specificity of the MicroVue DPD EIA?**

Information regarding the sensitivity, specificity, and precision for the MicroVue DPD EIA can be located on pages 4-5 of the MicoVue DPD EIA Package Insert:  
[http://www.quidel.com/sites/quidel.com/files/product/documents/8007\\_microvue\\_dpd\\_english\\_0.pdf](http://www.quidel.com/sites/quidel.com/files/product/documents/8007_microvue_dpd_english_0.pdf)

### **What is the target in this assay and how is it detected?**

The MicroVue DPD EIA targets urinary DPD crosslinks. The crosslinks are detected by a monoclonal antibody specific to DPD in an ELISA protocol. Please refer to pages 1-2 of the MicroVue DPD EIA Package Insert:  
[http://www.quidel.com/sites/quidel.com/files/product/documents/8007\\_microvue\\_dpd\\_english\\_0.pdf](http://www.quidel.com/sites/quidel.com/files/product/documents/8007_microvue_dpd_english_0.pdf)

### **What sample types can be used with this assay?**

#### **What are the storage conditions for samples, and requirements for sample handling?**

Information regarding proper sample type, storage, and handling can be found on page 3 of the MicroVue DPD EIA Package Insert:  
[http://www.quidel.com/sites/quidel.com/files/product/documents/8007\\_microvue\\_dpd\\_english\\_0.pdf](http://www.quidel.com/sites/quidel.com/files/product/documents/8007_microvue_dpd_english_0.pdf)

### **What is included in the kit?**

#### **What are the storage conditions for the kit and its components?**

A list of kit components and storage conditions can be found on pages 2 and 3 of the MicroVue DPD EIA Package Insert:  
[http://www.quidel.com/sites/quidel.com/files/product/documents/8007\\_microvue\\_dpd\\_english\\_0.pdf](http://www.quidel.com/sites/quidel.com/files/product/documents/8007_microvue_dpd_english_0.pdf)

### **How long is the Wash Solution stable for once it is prepared per the instructions in the Package Insert?**

#### **How long before use do I need to prepare the Working Substrate Solution?**

Specific information about preparation, storage, and stability of reagents can be found on page 3 of the MicroVue DPD EIA Package Insert:  
[http://www.quidel.com/sites/quidel.com/files/product/documents/8007\\_microvue\\_dpd\\_english\\_0.pdf](http://www.quidel.com/sites/quidel.com/files/product/documents/8007_microvue_dpd_english_0.pdf)

**What are the wavelength requirements for absorbance reading, and when should the absorbance be read?**

**Do I need to correct my results for variations in urine concentration by Creatinine?**

Instructions and guidelines for interpretation of results for the MicroVue DPD EIA can be found on pages 3-4 of the MicroVue DPD EIA Package Insert:

[http://www.quidel.com/sites/quidel.com/files/product/documents/8007\\_microvue\\_dpd\\_english\\_0.pdf](http://www.quidel.com/sites/quidel.com/files/product/documents/8007_microvue_dpd_english_0.pdf)

Refer to the Package Insert on our website at **quidel.com** for additional performance claims.

*\*Under federal and State law, it is the individual provider's responsibility to determine appropriate coding, charges and claims for a particular service.*

Policies regarding coding and payment levels can vary greatly from payer to payer and change over time. Quidel Corporation strongly recommends that providers contact their own regional payers to determine appropriate coding and charge or payment levels prior to submitting claims.

\*\*For state by state fee schedule go to [www.cms.gov](http://www.cms.gov)

0226BH1214D (01/15)