



Quidel Molecular HSV 1+2 Control Set *Positive and Negative*

QUIDEL



INTENDED USE

The Quidel Molecular HSV 1+2 Control Set is intended to be used as full process controls in molecular testing. This unassayed Control Set is for *in vitro* diagnostic use and consists of a highly purified, genomic DNA of HSV-1 and HSV-2 viruses. The negative control is HSV-1 and HSV-2 free matrix.

INSTRUCTIONS FOR USE

The Quidel Molecular HSV 1+2 controls should be used at the same volume as unknown samples in a validated molecular test methodology.

MATERIALS PROVIDED

Cat. #M109

Control Kit (3 vials) – Store at 2°C to 8°C

#	Component	Quantity
CONTROL +	HSV 1 Positive Control Part M5108	1 vial/kit 1.0 mL per vial
CONTROL +	HSV 2 Positive Control Part M5109	1 vial/kit 1.0 mL per vial
CONTROL -	Negative Control Part M5110	1 vial/kit 1.0 mL

WARNINGS AND PRECAUTIONS

- Treat all specimens, samples and controls as potentially infectious. Follow universal precautions when handling samples, this kit, and its contents.
- Wear suitable protective clothing, gloves, eye, and face protection when using this kit.
- Avoid contact with skin and eyes.
- Use micropipettes with an aerosol barrier or positive displacement tips for all procedures.
- Avoid microbial and cross contamination of the kit reagents. Follow Good Laboratory Procedures.
- Controls contain 0.09% Sodium Azide. Do not empty these controls into drains. Dispose of the containers in a safe way.
- Proper sample collection, storage, and transport are essential for correct results.¹
- Store assay reagents as indicated on their individual labels.
- SDS is available upon request or can be accessed on the product website.

STORAGE AND HANDLING OF KIT REAGENTS

Store the kit at 2°C to 8°C until the expiration date listed on the labels.

QUALITY CONTROL

The controls included in the Quidel Molecular HSV 1+2 Control Set are unassayed controls. Control ranges may vary based on testing method. A specific range should be determined for each desired testing method. Each

laboratory should establish their own Quality Control ranges and frequency of QC testing based on applicable local laws, regulations and standard good laboratory practice.

LIMITATIONS

- For *in vitro* diagnostic use.
- This control has been studied using a molecular testing methodology; it is not intended for use with other methodologies.

CUSTOMER AND TECHNICAL SUPPORT

To place an order or for technical support, please contact a Quidel Representative at 800.874.1517 (in the U.S.) or 858.552.1100 (outside the U.S.), Monday through Friday, from 8:00 a.m. to 5:00 p.m., Eastern Time. Orders may also be placed by fax at 740.592.9820. For e-mail support, contact custserv@quidel.com or technicalsupport@quidel.com. For services outside the U.S., please contact your local distributor. Additional information about Quidel, our products, and our distributors can be found on our website quidel.com.

REFERENCES

1. CLSI MM13-A: Guidance for the Collection, Transport, Preparation and Storage of Specimens for Molecular Methods 25:31, 2005.



M109 – Quidel Molecular HSV 1+2 Control Set



MDSS GmbH
Schiffgraben 41
30175 Hannover,
Germany



Quidel Corporation
2005 East State Street, Suite 100
Athens, OH 45701 USA
quidel.com

M109en v2014OCT22

GLOSSARY

REF

Catalogue number



CE mark of conformity

EC REP

Authorized representative
in the European Community

LOT

Batch code



Use by



Manufacturer



Temperature limitation



Intended use



Biological risks



Consult e-labeling
instructions for use

CONT

Contents/Contains

IVD

For *in vitro* diagnostic use

CONTROL +

Control positive

CONTROL -

Control negative
