

## SARS Antigen Quiz

For use with Sofia Q



Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the Sofia SARS Antigen FIA procedure using Sofia Q. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current Package Insert instructions and/or Quick Reference instructions included in the kit.

- 
1. **How often does the manufacturer recommend that external controls be run on the kit?**
- |                                   |  |
|-----------------------------------|--|
| a) Once every 20 tests            | c) Every first sample tested each day  |
| b) Once per each shipment of kits | d) Once per each untrained operator, once for each new shipment and/or new lot of kits |
- 
2. **At what temperature do the kits need to be stored?**
- |                                    |   |
|------------------------------------|---|
| a) 2°C to 8°C (refrigerated)       | c) 2°C to 30°C (refrigerated or room temperature) |
| b) 15°C to 30°C (room temperature) | d) -20°C (frozen)                                 |
- 
3. **Run both the positive and negative control swabs to meet the manufacturer recommendations.**
- a) True  
b) False
- 
4. **Which swab can be used to collect a patient specimen?**
- |                                     |  |
|-------------------------------------|--|
| a) A nasal swab supplied in the kit | c) A nylon flocked nasopharyngeal swab |
| b) A rayon throat swab              | d) Both A and C                        |
- 
5. **How long can a nasal swab sample be stored prior to testing?**
- |  |                               |
|--|-------------------------------|
| a) Samples should be tested as soon as possible          | c) Store samples indefinitely |
| b) Store up to 48 hours 2°C to 30°C (refrigerated or RT) | d) Both A and B               |
- 
6. **How much patient sample is added to the Test Cassette?**
- |  |              |
|--|--------------|
| a) One pipette contents (120 µL)       | c) One drop  |
| b) Entire contents of the Reagent Tube | d) Two drops |
- 
7. **In the WALK AWAY Mode, how long does the Test Cassette sit on the counter prior to inserting into Sofia Q?**
- a) Cassette is placed immediately inside analyzer  
b) Cassette sits on the counter for 15 minutes
- 
8. **In the READ NOW Mode, how long does the Test Cassette sit on the counter prior to inserting into Sofia Q?**
- a) Cassette is placed immediately inside analyzer  
b) Cassette sits on the counter for 15 minutes

## SARS Antigen Quiz

For use with Sofia Q



Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the Sofia SARS Antigen FIA procedure using Sofia Q. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current Package Insert instructions and/or Quick Reference instructions included in the kit.

- 
1. **How often does the manufacturer recommend that external controls be run on the kit?**
- a) Once every 20 tests
  - b) Once per each shipment of kits
  - c) Every first sample tested each day
  - d) Once per each untrained operator, once for each new shipment and/or new lot of kits
- 
2. **At what temperature do the kits need to be stored?**
- a) 2°C to 8°C (refrigerated)
  - b) 15°C to 30°C (room temperature)
  - c) 2°C to 30°C (refrigerated or room temperature)
  - d) -20°C (frozen)
- 
3. **Run both the positive and negative control swabs to meet the manufacturer recommendations.**
- a) True
  - b) False
- 
4. **Which swab can be used to collect a patient specimen?**
- a) A nasal swab supplied in the kit
  - b) A rayon throat swab
  - c) A nylon flocked nasopharyngeal swab
  - d) Both A and C
- 
5. **How long can a nasal swab sample be stored prior to testing?**
- a) Samples should be tested as soon as possible
  - b) Store up to 48 hours 2°C to 30°C (refrigerated or RT)
  - c) Store samples indefinitely
  - d) Both A and B
- 
6. **How much patient sample is added to the Test Cassette?**
- a) One pipette contents (120 µL)
  - b) Entire contents of the Reagent Tube
  - c) One drop
  - d) Two drops
- 
7. **In the WALK AWAY Mode, how long does the Test Cassette sit on the counter prior to inserting into Sofia Q?**
- a) Cassette is placed immediately inside analyzer
  - b) Cassette sits on the counter for 15 minutes
- 
8. **In the READ NOW Mode, how long does the Test Cassette sit on the counter prior to inserting into Sofia Q?**
- a) Cassette is placed immediately inside analyzer
  - b) Cassette sits on the counter for 15 minutes