



Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the Solana GAS Assay procedure. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current Package Insert instructions and/or Procedure Card provided on the Quidel website.

-
1. **What specimen types have been cleared for use with the assay?**
 - a) Blood
 - b) BAL
 - c) Urine
 - d) Throat and Eswab

 2. **Which of the following would be the appropriate amount of collected specimen to add to the Lysis Tube?**
 - a) Swab dip
 - b) 50 µL
 - c) 20 µL
 - d) Both a and b

 3. **What is the time frame and storage conditions for specimens prior to testing?**
 - a) Up to 2 days at room temperature
 - b) Up to 72 hours at room temperature
 - c) Up to 14 days at 2°C to 8°C
 - d) Must be tested immediately

 4. **Specimens in Lysis Tubes that cannot be immediately processed may be stored at what temperature and for how long?**
 - a) 20°C to 25°C or 2°C to 8°C for up to 24 hours
 - b) 20°C to 25°C or 2°C to 8°C for up to 48 hours
 - c) 20°C to 25°C or 2°C to 8°C for up to 5 days
 - d) 20°C to 25°C or 2°C to 8°C for up to 2 hours

 5. **Which of the following would be the proper way of mixing the Dilution Tube after adding the lysed sample?**
 - a) Inversion
 - b) Centrifugation
 - c) Vortex
 - d) Does not need mixing

 6. **Specimens in Dilution Tubes that cannot be processed right away may be stored at what temperature and for how long?**
 - a) 20°C to 25°C or 2°C to 8°C for up to 24 hours
 - b) 20°C to 25°C or 2°C to 8°C for up to 48 hours
 - c) 20°C to 25°C or 2°C to 8°C for up to 5 days
 - d) 20°C to 25°C or 2°C to 8°C for up to 2 hours

 7. **At what temperature do the Dilution, Lysis, and Reaction Tubes need to be stored?**
 - a) 20°C to 28°C
 - b) 2°C to 8°C
 - c) -20°C
 - d) -70°C

 8. **What is the volume of diluted specimen that needs to be added to the Reaction Tube?**
 - a) 15 µL
 - b) 145 µL
 - c) 50 µL
 - d) 125 µL

 9. **Which of the following may influence results from the Solana instrument?**
 - a) Vibrations
 - b) Highly intense light falling into the tube holder port
 - c) Ambient light
 - d) Both A and B

 10. **After Reaction Tubes are loaded into the Solana instrument, how long will it process before results are available?**
 - a) 25 minutes
 - b) 10 minutes
 - c) 1 hour
 - d) 2 hours

 11. **How many tests may be run at one time on the Solana instrument?**
 - a) 24 tests
 - b) 48 tests
 - c) 12 tests
 - d) 10 tests

 12. **What must the user do before moving Solana?**
 - a) Engage the electronic transport lock in the Main Menu
 - b) Turn off Solana
 - c) Engage the mechanical transport lock at the bottom of Solana
 - d) Both A and C

 13. **How often must you externally calibrate the Solana instrument?**
 - a) Once every 6 months
 - b) Solana does its own internal calibration
 - c) Every day the instrument is in use
 - d) With every run of the instrument

 14. **What form of technology is used to run the Solana GAS Assay?**
 - a) PCR
 - b) LAMP
 - c) HDA
 - d) Lateral flow immunoassay

 15. **You are able to access your data once you have navigated away from the results page by selecting which option from the main menu?**
 - a) New Test
 - b) System
 - c) Review Results
 - d) Results are not accessible once you navigate from the results page.

 16. **In addition to following Federal, State, and Local guidelines, how often should QC be run for the Solana GAS Assay?**
 - a) With each new lot and shipment
 - b) Once per shift
 - c) Each time the assay is performed
 - d) Once per day

 17. **What substances may be used to clean Solana?**
 - a) A cloth dampened with water
 - b) A cloth moistened with 70% alcohol
 - c) A cloth dipped in 1% bleach solution followed by water
 - d) All of the above

 18. **What are your options for making a test selection?**
 - a) Select a test from the drop-down menu
 - b) Manually type in the test name
 - c) Scan the barcode from the assay packaging
 - d) Both A and C

Circle the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the Solana GAS Assay procedure. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current Package Insert instructions and/or Procedure Card provided on the Quidel website.

1. What specimen types have been cleared for use with the assay?

- a) Blood
- b) BAL
- c) Urine
- d) Throat and Eswab

2. Which of the following would be the appropriate amount of collected specimen to add to the Lysis Tube?

- a) Swab dip
- b) 50 µL
- c) 20 µL
- d) Both a and b

3. What is the time frame and storage conditions for specimens prior to testing?

- a) Up to 2 days at room temperature
- b) Up to 72 hours at room temperature
- c) Up to 14 days at 2°C to 8°C
- d) Must be tested immediately

4. Specimens in Lysis Tubes that cannot be immediately processed may be stored at what temperature and for how long?

- a) 20°C to 25°C or 2°C to 8°C for up to 24 hours
- b) 20°C to 25°C or 2°C to 8°C for up to 48 hours
- c) 20°C to 25°C or 2°C to 8°C for up to 5 days
- d) 20°C to 25°C or 2°C to 8°C for up to 2 hours

5. Which of the following would be the proper way of mixing the Dilution Tube after adding the lysed sample?

- a) Inversion
- b) Centrifugation
- c) Vortex
- d) Does not need mixing

6. Specimens in Dilution Tubes that cannot be processed right away may be stored at what temperature and for how long?

- a) 20°C to 25°C or 2°C to 8°C for up to 24 hours
- b) 20°C to 25°C or 2°C to 8°C for up to 48 hours
- c) 20°C to 25°C or 2°C to 8°C for up to 5 days
- d) 20°C to 25°C or 2°C to 8°C for up to 2 hours

7. At what temperature do the Dilution, Lysis, and Reaction Tubes need to be stored?

- a) 20°C to 28°C
- b) 2°C to 8°C
- c) -20°C
- d) -70°C

8. What is the volume of diluted specimen that needs to be added to the Reaction Tube?

- a) 15 µL
- b) 145 µL
- c) 50 µL
- d) 125 µL

9. Which of the following may influence results from the Solana instrument?

- a) Vibrations
- b) Highly intense light falling into the tube holder port
- c) Ambient light
- d) Both A and B

10. After Reaction Tubes are loaded into the Solana instrument, how long will it process before results are available?

- a) 25 minutes
- b) 10 minutes
- c) 1 hour
- d) 2 hours

11. How many tests may be run at one time on the Solana instrument?

- a) 24 tests
- b) 48 tests
- c) 12 tests
- d) 10 tests

12. What must the user do before moving Solana?

- a) Engage the electronic transport lock in the Main Menu
- b) Turn off Solana
- c) Engage the mechanical transport lock at the bottom of Solana
- d) Both A and C

13. How often must you externally calibrate the Solana instrument?

- a) Once every 6 months
- b) Solana does its own internal calibration
- c) Every day the instrument is in use
- d) With every run of the instrument

14. What form of technology is used to run the Solana GAS Assay?

- a) PCR
- b) LAMP
- c) HDA
- d) Lateral flow immunoassay

15. You are able to access your data once you have navigated away from the results page by selecting which option from the main menu?

- a) New Test
- b) System
- c) Review Results
- d) Results are not accessible once you navigate from the results page.

16. In addition to following Federal, State, and Local guidelines, how often should QC be run for the Solana GAS Assay?

- a) With each new lot and shipment
- b) Once per shift
- c) Each time the assay is performed
- d) Once per day

17. What substances may be used to clean Solana?

- a) A cloth dampened with water
- b) A cloth moistened with 70% alcohol
- c) A cloth dipped in 1% bleach solution followed by water
- d) All of the above

18. What are your options for making a test selection?

- a) Select a test from the drop-down menu
- b) Manually type in the test name
- c) Scan the barcode from the assay packaging
- d) Both A and C