



Sofia®

Technical Bulletin

Set Barcode Scanner on Sofia to Default Settings

In order to bring the barcode scanner on the Sofia Analyzer back to the default settings, scan the following barcodes in order. When scanning the barcodes, please wait for about 1 second before scanning the next barcode.



Pause for one second. Reactivate trigger button on scanner.



Please contact Quidel Technical Support at 800.874.1517 (in the U.S.), 858.552.1100 (outside the U.S.) or technicalsupport@quidel.com if you have any questions regarding Sofia, or any Quidel product. Our hours of operation are Monday through Friday, 7:00 a.m. to 5:00 p.m. Pacific Time.

You may also visit our website at quidel.com for information on Quidel's line of Rapid Diagnostics, Molecular Diagnostics, Cell Culture and Specialty Diagnostics (Bone Health and Autoimmune & Complement). Other product information available on our website includes: CPT codes, CLSI procedure guides, MSDS, and Package Inserts.

0997M11211D-3 (03/14)